

new Prototype Detection System Based on the Beta-Gamma Coincidence Technique for Radioxenon Measurement.

Wednesday, June 21, 2023 11:22 AM (1 minute)

In this study, we develop a new radioxenon detection system based on beta-gamma coincidence to improve the sensitivity of detection in laboratory analysis. The designed system is a new prototype that is tested by injecting the ^{222}Rn and its daughters (^{214}Pb and ^{214}Bi) as a beta-gamma emitter and also ^{131}mXe gaseous sources, which are in the interest of CTBTO. Further, the system is calibrated and checked using a ^{166}mHo source. This system consists of a NaI (Tl) as a gamma, X ray radiation detector, and a Silicon detector for the beta or the conversion electron detection. Silicon is used for improving energy resolution and minimizing the potential for memory effect compared to the scintillator detector. The list-mode multi-parameter data acquisition system is used to set up coincidence parameters. The efficiency of two detectors and MDA are obtained.

E-mail

azimi.bme@gmail.com

Promotional text

A new prototype detection system based on the beta-gamma coincidence technique for radioxenon measurement.

Oral preference format

pre-recorded video

Primary author: Ms AZIMI, Sepideh Alsadat (Amirkabir University of Technology)

Co-authors: Mr AFARIDEH, Hossein (Amirkabir University of Technology); Ms FISEROVA, Lucie (NUVIA a.s.)

Presenter: Ms AZIMI, Sepideh Alsadat (Amirkabir University of Technology)

Session Classification: Lightning talks: P2.2, P3.2, P3.6

Track Classification: Theme 3. Monitoring and On-Site Inspection Technologies and Techniques: T3.2 Radionuclide Technologies and Applications