



ID: O3.1-814

Type: Oral

Field Calibration of Weak Motion Seismometers using Strong Motion Accelerometers as Reference

Tuesday 9 September 2025 14:15 (15 minutes)

Traceable calibration of broadband weak-motion seismometers is an ongoing challenge. In a specialized laboratory, verification of the response up to the upper corner frequency is achievable, but at a typical lower corner period accuracy is limited by parasitic tilts. Since such primary calibrations are not practical in the field, it has been proposed instead to use lab-calibrated reference seismometers in field calibrations, but then accuracy is limited at high frequency by the stiffness of the coupling between the sensors and other factors. There is, furthermore, the question of the stability of the response of the reference sensor with time and temperature.

We propose strong-motion accelerometers as an alternative reference. These instruments have several attractive characteristics. First, the response is flat to acceleration down to zero frequency. Second, some models have both a clip level high enough to measure the acceleration due to gravity and a noise floor low enough to resolve ocean microseisms. These properties mean absolute calibration of the reference can be performed on-site and relative calibration can be done in a side-by-side test. We demonstrate such a calibration and show how this could be part of routine traceable field calibration of IMS instrumentation.

E-mail

nicholas.ackerley@nrcan-rncan.gc.ca

In-person or online preference

Primary author: ACKERLEY, Nick (Canadian Hazards Information Service, Natural Resources Canada)

Co-author: IRVING, Callum (Canadian Hazards Information Service, Natural Resources Canada)

Presenter: ACKERLEY, Nick (Canadian Hazards Information Service, Natural Resources Canada)

Session Classification: O3.1 Seismic, Hydroacoustic and Infrasound Technologies and Applications

Track Classification: Theme 3. Monitoring and On-Site Inspection Technologies and Techniques: T3.1 Seismic, Hydroacoustic and Infrasound Technologies and Applications