

Erling Sandermann Olsen and Henrik Birch Carlsen
HBK-DPLA, Hottinger, Brüel & Kjær, Denmark

P3.1-887

- We have established a 10 mHz to 250 Hz calibration service
- Calibration of Brüel & Kjær microphones Type 4193-L-004
- Based on the work in Infra-AV project – and a lot of work before and after that
- We have determined the static pressure dependence of the microphones
 - This is necessary for the use of the calibrated microphone at any altitude
- There were a few interesting challenges during the development
- Come and hear more about the service and the work behind it in front of our poster

