

ID: P4.1-196

Type: e-Poster

## quality control check source for improved gain tracking and calibration

Friday, 2 July 2021 10:00 (15 minutes)

Quality control (QC) in beta-gamma coincidence systems that are used for radioxenon measurement is currently performed using a Cs-137 check source. Compton scattering in the scintillating plastic (polyvinyltoluene) of the beta detector creates a line of activity from the 667 keV gamma captured by the sodium iodide (NaI) across the beta dynamic range. Fitting the Compton scatter line provides a beta endpoint energy. The beta endpoint is used in combination with the 662 keV gamma peak to track the gain drift in the system before and after a sample measurement to ensure system stability during the count. Periodic QC checks are performed as well to monitor the long-term health of the system. By introducing another isotope to the check source, Ba-133, a second Compton scatter line can be generated in the coincidence counts that is well separated from the first. This would provide improved gain tracking, efficiency monitoring, and potential data on detector degradation. The multiple low-energy gamma lines of Ba-133, as compared to the 662 keV of Cs-137, would provide additional peaks for the energy and efficiency calibrations.

## **Promotional text**

This presentation provides a new method for improving the reliability and accuracy of nuclear test monitoring systems.

Primary author: Mr SLACK, Johnathan (Pacific Northwest National Laboratory (PNNL), Richland, WA, USA)

**Co-authors:** Mr MAYER, Michael (Pacific Northwest National Laboratory (PNNL), Richland, WA, USA); Mr COOPER, Matthew (Pacific Northwest National Laboratory (PNNL), Richland, WA, USA); Mr ELY, James (Pacific Northwest National Laboratory (PNNL), Richland, WA, USA); Mr FOXE, Michael (Pacific Northwest National Laboratory (PNNL), Richland, WA, USA);

Presenter: Mr SLACK, Johnathan (Pacific Northwest National Laboratory (PNNL), Richland, WA, USA)

Session Classification: T4.1 e-poster session

**Track Classification:** Theme 4. Performance Evaluation and Optimization: T4.1 - Performance Evaluation and Modelling of the Full Verification System and its Components