



ID: P2.4-078

Type: e-Poster

## results with INVAP STAX monitor

*Wednesday, June 30, 2021 11:45 AM (15 minutes)*

A general description of INVAP STAX monitor final status is shown, including calibration and software/hardware configuration.

First measurements results obtained in a real Noble Gas emission environment (MIPF plant at Ezeiza Atomic Center, CNEA, Buenos Aires, Argentina) are presented. Equipment measurement ranges, mainly focused to low flow rate and high activity concentration emissions, are discussed.

### Promotional text

Radioxenon accurately detection of high concentration emissions at MIPF such as EZEIZA (CAE, CNEA, ARGENTINA) is improved with the new STAX monitor manufactured by INVAP.

**Primary authors:** Mr FERNANDEZ BALDIS, Federico (INVAP S.E., Bariloche, Argentina); Mr NUNEZ, Mauro (INVAP S.E., Bariloche, Argentina); Mr PINO, Roman (INVAP S.E., Bariloche, Argentina); Mr ZAPATA, Andres (INVAP S.E., Bariloche, Argentina); Ms DI TADA, Mariana (CONICET, Instituto Tecnológico de Buenos Aires (ITBA), Buenos Aires, Argentina); Mr NASSIF, Eduardo (INVAP S.E., Bariloche, Argentina); Mr SAGARZAZU, Ricardo (INVAP S.E., Bariloche, Argentina); Mr BOCCOLI, Horacio (INVAP S.E., Bariloche, Argentina); Mr CARRANZA, Eduardo (National Atomic Energy Commission (Comisión Nacional de Energía Atómica), Argentina)

**Presenter:** Mr FERNANDEZ BALDIS, Federico (INVAP S.E., Bariloche, Argentina)

**Session Classification:** T2.4 e-poster session

**Track Classification:** Theme 2. Events and Nuclear Test Sites: T2.4 - Atmospheric and Subsurface Radionuclide Background and Dispersion